

Utilizzo di elettrodi di terze parti con defibrillatori TECNO-GAZ

Con la presente, Noi, TECNO-GAZ S.p.A., desideriamo informare tutti i partner del settore che l'uso di piastre di defibrillazione di terze parti per i nostri modelli di TECNO HEART PLUS e TECNO HEART SMART non è autorizzato ed è severamente sconsigliato.

Solo le piastre originali fornite da TECNO-GAZ S.p.A. possono garantire prestazioni accurate del dispositivo.

L'uso di piastre non autorizzate può causare un funzionamento improprio del DAE e fornire misurazioni errate.

TECNO-GAZ S.p.A. non può assumersi alcuna responsabilità per eventuali danni causati dall'uso di elettrodi non autorizzati con le nostre apparecchiature.

Use of 3rd party pads with TECNO-GAZ AEDs

Hereby, we, TECNO-GAZ S.p.A., would like to inform all channel partners, that the use of 3rd party defibrillation pads for our AED models TECNO HEART PLUS e TECNO HEART SMART is not authorized and strictly warned against.

Only original pads supplied by TECNO-GAZ S.p.A., can guarantee the accurate performance of the device.

The use of unauthorized pads may cause the AED to operate improperly and provide false measurements.

TECNO-GAZ S.p.A. cannot take any responsibility for potential damages caused by the use of unauthorized pads with our equipment.

Sala Baganza (PR) 29-09-2023


Franco Ravanetti
Amministratore Delegato - CEO

Firma autografa sostituita a mezzo stampa, ai sensi dell'art. 3, comma 2, del D.lgs. 39/93"